

## **E&E** Certification Record

Listing#: E116062 Report #:132907

Original Certification Date: November 19, 2024

This Certification is issued to: Amico Mobility Solutions 122B East Beaver Creek Rd. Richmond Hill, ON, L4B 1G6



Stating that the product(s):

Medical Pendant and Console,

Models: LP-L-xxxx-N-E, LP-L-xxxx-R-E, LP-L-PLP-xxxx-R-E, LP-H-NNNN, LP-H-NNNN-R-E, LP-V-xxxx -N-E-90, and LP-V-TM-xxxx-N-E-90

Product Rating(s):

120V or 220V or 240Vac; 20A or 15A or 13A; 50/60Hz

Duty Cycle: On: 1 min. Off: 9 min.

Class I

Achieved Certification to the following standard(s):

ANSI/AAMI ES60601-1: 2005/(R)2012 - Medical electrical equipment — Part 1: General

requirements for basic safety and essential performance, and

CSA-C22.2 No. 60601-1: 2014 - Medical electrical equipment — Part 1: General requirements for basic safety and essential performance

Richard Johnson Certification Officer.

Eurofins Electrical and Electronic Testing North America, Inc.

All changes proposed in the previously identified product that affects the above information must be submitted to Eurofins for evaluation prior to implementation to assure continued NRTL Certification status. The covered product(s) shall be subject to follow-up inspections to ensure that the Certified product(s) are identical to the product sample evaluated by Eurofins E&E NA and that all responsibilities are being fulfilled as specified in the Applicants' Responsibility section of the Certification Report. The Applicant named above has been authorized Eurofins E&E NA to represent he product(s) listed in this record as "MET Certified" and to mark this/these product(s) according to the terms and conditions of the Eurofins E&E NA Applicant Contract, Listing Reports, and the applicable agreements. Only the product(s) bearing the MET Mark and under a follow-up service are considered to be included in this Certification program. This certification has been granted under a System 3 program as defined in ISO/IEC 17067.



Eurofins E&E North America, Inc. is accredited by OSHA and the Standards Council of Canada

