Spacelabs Qube VESA Adaptor Plate
Quick Installation and Maintenance Instructions

MON-VESA-ADA-DRP-S

Medical Facility Responsibility

Preventive maintenance checks must be performed regularly to maintain the quality and performance of this product. Any parts that may be broken, missing, worn, distorted or contaminated in any way should not be used and all affected parts should be replaced immediately. Should the necessity of any repair be suspected; please contact your local Amico distributor.

WARNING: It is the responsibility of the end user to ensure all aspects of installation are covered.

Installation Tools

#2 Phillips Head Screwdriver

Installation Reference

MON-VESA-ADA-DRP-(-S)

Installation to Device (Spacelabs Qube/Qube Mini)

1. Lay the *MONITOR* on a padded surface and align parts as shown.
2. Insert 4x *H-PHMS-M4-10* through the *X-MON-VESA-ADA-DRP-S* and into the *MONITOR*.



WARNING: Insert and tighten M4 *SCREWS* in a cross pattern.

Installation to Mounting Platform (Amico #2 Head and - S Head)

1. Pull out *X-MON-PLUNPIN* from *X-DROP-SHLF* and rotate 90° to lock in outer position. *
2. Ensure 2x *H-PHSCS-1420-12* are loosened and do not protrude from the front surface of the *X-DROP-SHLF*. **



Spacelabs Qube VESA Adaptor Plate
Quick Installation and Maintenance Instructions

MON-VESA-ADA-DRP-S

3. With *MONITOR* mounted on *X-MON-VESA-ADA-DRP-S* (not shown), slide *X-MON-VESA-ADA-DRP-S* into *X-DROP-SHLF* by lining up 2x *X-MON-BLK-STDOF* with the 2x *CUTOUTS* on *X-DROP-SHLF* * so that 2x *X-MON-BLK-STDOF* and *X-MON-VESA-ADA-DRP-S* sandwich the thinner portion of the 2x *CUTOUTS* on *X-DROP-SHLF*. **
4. Rotate *X-MON-PLUNPIN* 90° to release it through the corresponding hole in *X-MON-VESA-ADA-DRP-S* ***, and hand tighten 2x *H-PHSCS-1420-12* to secure *X-MON-VESA-ADA-DRP-S* in *X-DROP-SHLF*.



⚠ **WARNING:** Insert and tighten M4 *SCREWS* in a cross pattern.

Troubleshooting Guide

Symptom	Possible Cause	Solution
Monitor is not secure.	Loose mounting to bracket.	Re-align parts and refer to instructions (Installation to Device).
Monitor is secure but bracket is not.	Loose mounting to mounting platform.	Re-align parts and refer to instructions (Installation to Mounting Platform).

Maintenance

It is recommended that once every (3) months, all locks, bolts, and screws are visually inspected for signs of product wear and tear or damage. Check the product for looseness in the mount or orientation.

Warranty

During the term of warranty: Within the first twelve (12) months from the date of shipment, Amico Accessories will repair or replace any part which is proven to be defective at no cost. After the twelve (12) month period, Amico Accessories will send the parts to the customer free of charge, however shipping and installation will be borne by the customer.

The warranty is valid only when the product has been properly installed according to Amico Accessories specifications, used in a normal manner, and serviced according to factory recommendations. It does not cover failures due to damage which occur in shipments or failures which resulted from accidents, misuse, abuse, neglect, mishandling, alteration, misapplication or damage that may be attributable to Force Majeure.

AMICO ACCESSORIES DOES NOT HONOR VERBAL STATEMENTS CONCERNING THE WARRANTY.

The distributor and/or dealer are not sanctioned to create verbal warranties about the product described in this agreement. Any statements will not be honored or be made part of the agreement of sale. This document is the final complete and exclusive terms of the agreement.

THIS WARRANTY IS INCLUSIVE AND REPLACES ALL OTHER WARRANTIES.

Amico Accessories shall not, under any circumstances be liable for incidental or consequential damages including, but not limited to, profit, loss of sales or injuries to person(s) or property.

Correction of non-compliance as noted above will result in completion of all liabilities of Amico Accessories whether based on agreement, neglect or changed materials, designs or specifications without notice.

All claims for warranty must first be approved by Amico Accessories Customer Service Department: (info@amico-accessories.com or 1-877-264-2697). A valid Return Goods Authorization number must be obtained from Amico Accessories prior to commencement of any warranty claim.