

# NuLook FlexWall Surface Mounted



## General Specifications

The **NuLook FlexWall Surface Mounted System** shall be manufactured by Amico Corporation, in accordance with job specific submittal package. Shop drawing package includes gas termination and electrical connections. The following is a general specification, and components listed may not be present or required in final product.

The **NuLook FlexWall Surface Mounted System** is listed by Underwriters Laboratories Inc. (UL) in Canada and the United States of America. Amico Corporation manufactures all products in accordance with the CSA Z7396.1-06 standard, National Fire Protection Association (NFPA), NFPA-99c (current edition), National Electric Code (NEC), and the Canadian Electrical Code (CEC).

The installation contractor is responsible for compliance with all local, state, and federal codes applicable to the installation of medical gas and electrical systems.

### The NuLook FlexWall Surface Mounted System is Made Up of Several Integral Components:

- Central section with removable 3D wrap panels for bed area. This component is located between the vertical sections.
- Vertical sections where the services are located. These sections are installed on the finished wall. Electrical and medical gas terminations are located above the ceiling. Decorative side vertical sections with removable 3D wrap panels are available and are located to the left or right of the vertical sections.
- All panel sections are installed on the finished wall and have optional horizontal/vertical reveals or rails.
- The fascia of the panels is 1/2" Medium-density fiberboard with 3D wrap finish (typical) and 5/8" top access panel (typical). All fascia panels shall be removable for access and easy installation or maintenance.

## Materials and Construction

### Medical Gas Outlets

Amico Corporation has a variety of options available including DISS, Ohmeda/Medaes, Puritan-Bennett, Chemetron, & Oxequip/Medstar in different finishes.

### Medical Gas Manifold

All medical gas outlets come with a DISS flexhose that connects to the gas manifold within the Amico headwall. Amico Gas manifold locations are indicated on the final shop drawings. They ship loose and are mounted and connected to the source on site by others. Gas manifolds can be sized to allow for future expandability.

### Electrical Wiring

Wire for standard and critical branch power circuits shall be #12 type THHN stranded copper wire, 600 volts, with heat resistant thermoplastic insulation for hot (black) and neutral (white). Grounds shall be #10 THHN stranded copper wire (green) for all electrical circuits pre-wired by Amico.

### Grounding and Bonding

All ground conductors shall be installed in headwall or raceway. Each power receptacle shall have a ground conductor connected to a grounding screw. Where electrical terminations are located inside the headwall, a grounding bus will be installed for each type of power, and shall be installed as to insure grounding for the complete power system.

### Low Voltage Data Provisions

Shall be terminated with conduit stub-up in designated location. A pull string extending from conduit stub-up to service provision will be provided. Telephone jack, nurse call patient station device, data port, or other low voltage device shall be furnished, installed, and wired by others.

### Electrical Devices

This shall be Hospital Grade 15 or 20 amp, 120 or 277 volts, UL listed and shown on submittal shop drawings. Electrical contractor shall ensure compatibility of plug on accessory equipment to be used with these devices:

Duplex receptacles shall be NEMA style 5-15R or 5-20R, in ivory color for use on normal power circuits, and in red color for use on critical branch power circuits.

## Warranty

Amico Corporation warrants its Architectural Products to be free from material defects and workmanship for a period of ten (10) years from the date of shipment.

## Distributed By: